

State of Delaware Department of Correction

DATE: May 31, 2011

TO: ALL OFFERRS

FROM: Paul Giery
Purchasing Services Administrator

SUBJECT: ADDENDUM TO REQUEST FOR PROPOSAL (RFP) - CONTRACT NO.
DOC1102_URINALYSIS
OFFENDER URINALYSIS SCREEN TESTING

ADDENDUM # 5

This addendum is to advise all vendors of the below changes to appendix "A", Section 1, Testing Methodologies listed below. All other terms and conditions remain the same.

Testing Methodologies:

Initial screening shall be done by Enzyme Multiplied Immunoassay Technique (EMIT). The immunoassays will be grouped as follows and the cutoff levels will be as indicated:

Drug Screen A	ng/ml	Drug Screen B	ng/ml	Drug Screen C	Ng/m
Cannabinoid	50	Barbiturates	200	Synthetic	
Cocaine Metabolite	300	Methaqualone	300	Cannabinoids	
Opiates	300	Propoxyphene	300	(K2/Spice)	
Benzodiazepines	200	Phencyclidine	25		
Methadone	300	(PCP)			
Amphetamines	1000				

If you have any questions, please contact me at (302) 857-5262, Paul.Giery@state.de.us